



**ADWA Hungary Kft.**  
**Alsókikötő sor 11. – 6726 Szeged - Hungary**  
**Tel.: +36 62 317 878 - Fax: +36 62 550 610**  
**E-mail : sales@adwainstruments.com**

---



## **DECLARATION OF CONFORMITY**

**We**  
**ADWA Hungary Kft.**  
**Alsókikötő sor 11.**  
**6726 Szeged**  
**Hungary**

Hereby certify that the waterproof pH / Temp pocket tester:

**AD12**

Has been tested and found to be in compliance with EMC Directive 89/336/EEC and Low Voltage Directive 73/23/EEC according to the following applicable normative:

<b>EN 50082-1:</b>	<b>Electromagnetic Compatibility – Generic Immunity Standard</b>
<b>IEC 61000-4-2:</b>	<b>Electrostatic Discharge</b>
<b>IEC 61000-4-3:</b>	<b>RF Radiated</b>
<b>EN 50081-1:</b>	<b>Electromagnetic Compatibility – Generic Emission Standard</b>
<b>EN 55022:</b>	<b>Radiated, Class B</b>
<b>EN 61010-1:</b>	<b>Safety requirements for electrical equipment for measurement, control and laboratory used.</b>